

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2026-09-24 14:16:02
Date of Next Testing: 2028-09-23 14:16:02
Product Name: **Anthracene**
CAS Number: 120-12-7
Grades & Purity: Melting point standard
Lot #: I2616268
SKU #: **A119534**
Version: **1**
Analysis #: **778424**
Storage Temperature: Room temperature

Parameter	Limit Values	Result
Capillary melting point(0.20°C/min)	216.22-216.44 (°C)	216.27000 °C
Capillary melting point(1.0°C/min)	216.72-217.12 (°C)	216.87000 °C
Capillary melting point((C°)F = 1,ΔT = 0))	215.83-215.93 (°C)	215.83000 °C



John Su
QA & QC Manager