

## Certificate of Analysis

Date of Testing: 2023-12-20 15:28:21  
Date of Next Testing: 2025-12-19 15:28:21  
Product Name: **Anthracene**  
CAS Number: 120-12-7  
Specifications & Purity: Melting point standard  
Lot #: L2301077  
SKU #: **A119534**  
**Analysis:** **299193**  
Storage Temperature: Room temperature

Parameter	Limit Values	Result
Capillary melting point(0.20°C/min)	216.22-216.44 (°C)	216.31 °C
Capillary melting point(1.0°C/min)	216.72-217.12 (°C)	216.91 °C
Capillary melting point((C°)F = 1,ΔT = 0))	215.83-215.93 (°C)	215.87000 °C



John Su  
QA & QC Manager