

## Certificate of Analysis

Date of Testing: 2025-03-18 17:19:44

Date of Next Testing: 2026-03-18 17:19:44

Product Name: **Atrazine Standard**

CAS Number: 1912-24-9

Specifications & Purity: 1000ug/ml in Acetone

Lot #: C2512203

SKU #: **A128222**

Version: **1**

Analysis #: **486036**

Storage Temperature: Room temperature

| Parameter                         | Limit Values    | Result       |
|-----------------------------------|-----------------|--------------|
| Concentration                     | 900-1100 (mg/L) | 1000.00 mg/L |
| Related expanded uncertainty(k=2) | 0-4 (%)         | 4.00000 %    |
| Solvent(Acetone)                  | Conform         | Conform      |



John Su  
QA & QC Manager