

ALADDIN SCIENTIFIC CORPORATION  
14078 Meridian Parkway, Riverside, CA. 92518

### Certificate of Analysis

Date of Testing: 2026-04-17 16:16:13  
Date of Next Testing: 2027-04-17 16:16:13  
Product Name: **Glucose Colorimetric Detection Kit t (GOD-POD Microplate Method)**  
CAS Number:  
Specifications & Purity: BioReagent  
Lot #: D2617620  
SKU #: **A1371397**  
Version **1**  
Analysis #: **695475**  
Storage Temperature: Store at 2-8°C, Protected from light, Room temperature, Store at -20°C, Avoid repeated freezing and thawing

Parameter	Limit Values	Result
Appearance(A1371397A)	Slight yellow, transparent, odorless and no precipitation liquid	Slight yellow, transparent, odorless and no precipitation liquid
Appearance(A1371397B)	Slight yellow, transparent, odorless, no precipitation and shake with foam liquid	Slight yellow, transparent, odorless, no precipitation and shake with foam liquid
Appearance(A1371397C)	Colorless, transparent, odorless and no precipitation liquid	Colorless, transparent, odorless and no precipitation liquid
Appearance(A1371397D)	Colorless, transparent, odorless and no precipitation liquid	Colorless, transparent, odorless and no precipitation liquid
Seal Integrity Testing	Conform	Conform
Density:0.98~1.02(g/mL) (A1371397A)	Conform	Conform
Density:0.97~1.03(g/mL) (A1371397B)	Conform	Conform
Performance(Glucose Assay)	Conform	Conform



QualityAssurance@aladdinsci.com

---

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

A handwritten signature in black ink that reads "John Su". The signature is written in a cursive, flowing style.

John Su  
QA & QC Manager