



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

## Certificate of Analysis

Date of Testing: 2026-07-09 15:45:13

Date of Next Testing: 2028-01-05 15:45:13

Product Name: **AP Microwell ECL substrate**

CAS Number:

Specifications & Purity:

Lot #: A2404930

SKU #: **A615645**

Version **1**

**Analysis #:** **735721**

Storage Temperature: Store at 2-8°C, Protected from light

| Parameter                       | Limit Values | Result         |
|---------------------------------|--------------|----------------|
| pH                              | 9.3-10.1     | 9.35800        |
| Background RLU                  | 0-2000       | 360.00000      |
| Signal RLU                      | ≥100000000   | 15406000.00000 |
| Appearance(light yellow liquid) | pass         | Pass           |

John Su  
QA & QC Manager