



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2024-01-29 11:20:49

Date of Next Testing: 2025-07-27 11:20:49

Product Name: **AP Microwell ECL substrate**

CAS Number:

Specifications & Purity:

Lot #: A2417172

SKU #: A615645

Version: 1

Analysis #: 309179

Storage Temperature: Store at 2-8°C, Protected from light

Parameter	Limit Values	Result
pH	9.3-10.1	9.50000
Background RLU	0-1000	186.00000
Signal RLU	≥5000000	5347269.00000
Appearance(Pale yellowish-green liquid)	Pass	Pass

John Su
QA & QC Manager