

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

## Certificate of Analysis

Date of Testing: 2025-05-14 16:54:36  
Date of Next Testing: 2026-11-10 16:54:36  
Product Name: **AP Blotting ECL substrate**  
CAS Number:  
Specifications & Purity:  
Lot #: A2425174  
SKU #: **A615662**  
Version: **1**  
Analysis #: **509341**  
Storage Temperature: Store at 2-8°C, Protected from light

Parameter	Limit Values	Result
PH	9.3-10.1	9.30
Background RLU	0-1000	266.00000
Signal RLU	≥5000000	5563365.00000
Appearance(Pale yellowish-green liquid)	Pass	Pass



John Su  
QA & QC Manager