

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-05-14 16:57:07
Date of Next Testing: 2026-11-10 16:57:07
Product Name: **AP Blotting ECL substrate**
CAS Number:
Specifications & Purity:
Lot #: A2425176
SKU #: **A615662**
Version: **1**
Analysis #: **509343**
Storage Temperature: Store at 2-8°C, Protected from light

Parameter	Limit Values	Result
PH	9.3-10.1	9.68
Background RLU	0-1000	261.98000
Signal RLU	≥10000000	11746562.00000
Appearance(Pale yellowish-green liquid)	Pass	Pass



John Su
QA & QC Manager