

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-11-25 18:55:21
Date of Next Testing: 2027-11-25 18:55:21
Product Name: **Chloride Standard**
CAS Number:
Specifications & Purity: 500µg/ml in Water (20°C)
Lot #: K2520301
SKU #: **C117300**
Version: **1**
Analysis #: **613186**

Parameter	Limit Values	Result
Concentration	495-505 (µg/ml)	500.00 µg/ml
Related expanded uncertainty(k=2)	0-4 (%)	1.00000 %
Appearance (C117300)	Colorless, transparent liquid	Colorless, transparent liquid
Solvent(H ₂ O)	Conform	Conform



John Su
QA & QC Manager