

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2026-06-18 16:43:28
 Date of Next Testing: 2027-06-18 16:43:28
 Product Name: **Cellulase (CL) Activity Assay Kit (DNS, Colorimetric Method)**
 CAS Number:
 Specifications & Purity: BioReagent
 Lot #: F2618030
 SKU #: **C1507307**
 Version: **1**
 Analysis #: **729328**
 Storage Temperature: Store at 2-8°C, Protected from light, Room temperature

Parameter	Limit Values	Result
Seal Integrity Testing	Conform	Conform
Appearance (C1507307A)	Colorless and transparent Solution, no precipitate	Colorless and transparent Solution, no precipitate
Odor (C1507307A)	Odorless	Odorless
Density (C1507307A)	Conform	Conform
(0.99 - 1.05 g/mL)		
Appearance (C1507307B)	Colorless and transparent Solution, no precipitate	Colorless and transparent Solution, no precipitate
Odor (C1507307B)	Odorless	Odorless
Density (C1507307B)	Conform	Conform
(1.00 - 1.05 g/mL)		
pH (C1507307B) (5.5-6.5)	Conform	Conform
Appearance (C1507307C)	Colorless or milky white and semi-transparent solution, no precipitate	Colorless or milky white and semi-transparent solution, no precipitate
Odor (C1507307C)	Specific Odor	Specific Odor
Density (C1507307C)	Conform	Conform
(1.00 - 1.05 g/mL)		
Appearance (C1507307D)	Orange and transparent solution, no precipitate	Orange and transparent solution, no precipitate
Odor (C1507307D)	Odorless	Odorless
Density (C1507307D)	Conform	Conform
(1.06 - 1.15 g/mL)		



QualityAssurance@aladdinsci.com

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

A handwritten signature in black ink that reads "John Su".

John Su
QA & QC Manager