

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2026-04-21 17:58:52

Date of Next Testing: 2029-04-20 17:58:52

Product Name: **Blood Coagulant (Enzyme-Containing)**

CAS Number:

Specifications & Purity:

BioReagent

Lot #: ZJ26F0434462

SKU #: **C1520197**

Version: **1**

Analysis #: 032997

Parameter	Limit Value	Result
pH	6.0-7.5	6.48100
Seal Integrity Testing	Conform	Conform
Appearance(C1520197)	Pale yellow turbid liquid; pale yellow precipitate or stratification may form upon standing; no visible impurities	Pale yellow turbid liquid; pale yellow precipitate or stratification may form upon standing; no visible impurities
Suspension Stability (No precipitation or stratification within 5 minutes)	Conform	Conform
Coagulation Time (5-10 min at RT.)	Conform	Conform
Microbial Count (≤ 10 CFU/ml)	Not Detected	Not Detected
Coagulation Promotion Effect (No fibrin precipitation; No hemolysis; Serum precipitation rate $\geq 50\%$.)	Conform	Conform



Sophia Cheng
Quality Assurance