

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-09-29 10:27:51

Date of Next Testing: 2029-09-28 10:27:51

Product Name: **Colour Coded Buffers(pH 10.00 ±0.01 Blue)**

CAS Number:

Specifications & Purity: Buffer Solution pH 10.00 (Blue) ±0.01

Lot #: I2524608

SKU # **C196915**

Version **0**

Analysis #: **584797**

Storage Temperature: Room temperature

Parameter	Limit Values	Result
pH(25 °C)	9.99-10.01	10.00000
Appearance (C196915)	Blue Liquid	Blue Liquid



John Su
QA & QC Manager