



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2023-04-19 15:14:25
Date of Next Testing: 2027-04-18 15:14:25
Product Name: **Conductivity Standards**
CAS Number:
Specifications & Purity: 84μS/cm @ 25°C
Lot #: D2314908
SKU #: **C196990**
Version: **1**
Analysis #: **190818**
Storage Temperature: Room temperature

Parameter	Limit Values	Result
Electric Conductivity(25°C)	82.32-85.68 (μS/cm)	84.00000 μS/cm

John Su
QA & QC Manager