



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2022-08-26 18:49:18
Date of Next Testing: 2026-08-25 18:49:18
Product Name: **Conductivity Standards**
CAS Number:
Specifications & Purity: 84μS/cm @ 25°C
Lot #: H2217587
SKU #: **C196990**
Version: **1**
Analysis #: **121356**
Storage Temperature: Room temperature

Parameter	Limit Values	Result
Electric Conductivity(25°C)	82.32-85.68 (μS/cm)	84.00000 μS/cm

John Su
QA & QC Manager