



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2023-09-06 16:23:02
Date of Next Testing: 2027-09-05 16:23:02
Product Name: **Conductivity Standards**
CAS Number:
Specifications & Purity: 84 μ S/cm @ 25°C
Lot #: I2305301
SKU #: **C196990**
Version: **1**
Analysis #: **265386**
Storage Temperature: Room temperature

Parameter	Limit Values	Result
Electric Conductivity(25°C)	82.32-85.68 (μ S/cm)	84.00000 μ S/cm

John Su
QA & QC Manager