

ALADDIN SCIENTIFIC CORPORATION
14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2026-01-16 14:55:59
Date of Next Testing: 2030-01-15 14:55:59
Product Name: **Citrate buffer**
CAS Number:
Specifications & Purity: 0.5M,pH3.5
Lot #: F2207002
SKU #: **C301547**
Version: **0**
Analysis #: **641365**

Parameter	Limit Values	Result
pH(25 °C)	3.3-3.7	3.55
Concentration	0.48-0.52 (mol/L)	0.50 mol/L
Appearance (C301547)	Colorless, transparent liquid	Colorless, transparent liquid



John Su
QA & QC Manager