

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

## Certificate of Analysis

Date of Testing: 2026-05-24 23:24:35

Date of Next Testing: 2027-05-24 23:24:35

Product Name: **Fecal Occult Blood Test Kit (OX-Toluidine Method)**

CAS Number:

Specifications &amp; Purity:

BioReagent

Lot #: ZJ25F0724833

SKU #: **D1354183**

Version: **1**

Analysis #: 036400

| Parameter                          | Limit Value                                                            | Result                                                                 |
|------------------------------------|------------------------------------------------------------------------|------------------------------------------------------------------------|
| Appearance(D1354183A)              | Gray transparent and no precipitation liquid with pungent smell liquid | Gray transparent and no precipitation liquid with pungent smell liquid |
| Appearance(D1354183B)              | Colorless, transparent, odorless and no precipitation liquid           | Colorless, transparent, odorless and no precipitation liquid           |
| Seal Integrity Testing             | Conform                                                                | Conform                                                                |
| Performance                        | Conform                                                                | Conform                                                                |
| Density(0.90-0.96 g/ml)(D1354183A) | Conform                                                                | Conform                                                                |
| Density(0.99-1.05 g/ml)(D1354183B) | Conform                                                                | Conform                                                                |



Sophia Cheng  
Quality Assurance