

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2026-06-24 12:17:00

Date of Next Testing: 2026-12-21 12:17:00

Product Name: **Mouse Hexosaminidase B Beta (β-Hex) ELISA Kit**

CAS Number:

Specifications & Purity: BioReagent, for enzyme immunoassay

Lot #: ZJ26F0636705

SKU #: **EJ1512586**

Version: **1**

Analysis #: 036716

Parameter	Limit Value	Result
Outer box and label: Shall be complete and standardized; batch number, validity period and specification shall be correct; no damage to outer box	Conform	Conform
Pre-coated plate: Aluminum foil bag shall be sealed with desiccant inside, no damage or air leakage	Conform	Conform
Liquids in dropping bottles: Including washing solution, sample dilution buffer, chromogenic solution, stop solution, etc.; shall be clear without suspended solids and precipitation	Conform	Conform
Kit components: Complete components with intact packaging	Conform	Conform
Liquid filling volume: The filling volume of each reagent in the components shall not be lower than the indicated content	Conform	Conform
Positive coincidence rate: Test 8 samples of positive serum respectively, all shall be positive	Conform	Conform
Negative coincidence rate: Test 8 samples of blank control respectively, all shall be negative	Conform	Conform
Inter-batch variation: The coefficient of variation of test results of kits from different batches shall be ≤ 10%	Conform	Conform

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Parameter	Limit Value	Result
Intra-batch variation: The coefficient of variation of test results of kits from the same batch shall be $\leq 10\%$	Conform	Conform
Stability: After destructive test at 37°C, test 8 tubes of standard substance, all shall show positive results	Conform	Conform



Sophia Cheng
Quality Assurance