

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-04-25 13:50:00
Date of Next Testing: 2028-04-24 13:50:00
Product Name: **HRP Stabilizer**
CAS Number:
Specifications & Purity:
Lot #: D2522312
SKU #: **H615758**
Version: **1**
Analysis #: **498860**
Storage Temperature: Store at 2-8°C

Parameter	Limit Values	Result
PH	7.0-7.4	7.20
Functional stability	90-100 (%)	98.90000 %
Appearance(light yellow liquid)	pass	Pass



John Su
QA & QC Manager