

ALADDIN SCIENTIFIC CORPORATION  
14078 Meridian Parkway, Riverside, CA. 92518

## Certificate of Analysis

Date of Testing: 2023-02-06 13:43:20  
Date of Next Testing: 2025-01-26 13:43:20  
Product Name: **Fluoride Ion Selective Electrode Solutions**  
CAS Number:  
Specifications & Purity: 1000ppm Standard  
Lot #: G1922229  
SKU #: **I123834**  
**Analysis: 183325**

Parameter	Limit Values	Result
Concentration	990-1010 (ppm)	999.00 ppm
Appearance (I123834)	Colorless Liquid	Colorless liquid



John Su  
QA & QC Manager