

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2026-04-08 15:54:39

Date of Next Testing: 2027-04-08 15:54:39

Product Name: **IPL-41 Insect Medium (1X)**

CAS Number:

Specifications & Purity: Bioactive, Sterile-Filtered, Sterile, for cell culture, suitable for insect cell culture

Lot #: ZJ25F0622913

SKU #: **I778237-500ml**

Version: **1**

Analysis #: 033745

Parameter	Limit Value	Result
pH	6.0-6.4	6.20000
Osmolality	300-345(mOsm/kg)	320.00000 mOsm/kg
Appearance(I778237)	Orange-brown and transparent liquid	Orange-brown and transparent liquid
Seal Integrity Testing	Conform	Conform
Endotoxin Level ≤ 0.25 EU/mL	Conform	Conform
Sterility test	Negative	Negative
Cell growth	Conform	Conform



Sophia Cheng
Quality Assurance