

Certificate of Analysis

Date of Testing: 2024-09-03 17:42:27

Date of Next Testing: 2028-09-02 17:42:27

Product Name: **analyte for coulometric Karl Fischer titration. Suitable for cells with and without diaphragm**

CAS Number:

Specifications & Purity: Used a diaphragm-type cathode electrolytic fluid

Lot #: H2430469

SKU #: **K119809**

Version: **2**

Analysis #: **394974**

Parameter	Limit Values	Result
Adjusted value(0.1 µl H ₂ O)	97-103 (ppm)	103.00000 ppm
Sensitivity(color changes rapidly and visibly)	pass	Pass



John Su
QA & QC Manager