

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2023-11-23 08:49:41
Date of Next Testing: 2025-11-12 08:49:41
Product Name: **Coulometric Karl Fischer Reagent (with diaphragm)**
CAS Number:
Specifications & Purity: Pyridine free, 1.5gH₂O/100ml, Used as anode electrolyte
Lot #: K2321433
SKU #: **K293428**
Analysis: 292898
Storage Temperature: Desiccated

Parameter	Limit Values	Result
PH	4-6	5.89
Sensitivity	Pass	Pass
Correction value(0.1ul H ₂ O)(97-103ppm)	Pass	Pass
Appearance(K293428)	Red brown liquid	Red brown liquid



John Su
QA & QC Manager