

ALADDIN SCIENTIFIC CORPORATION  
14078 Meridian Parkway, Riverside, CA. 92518

### Certificate of Analysis

Date of Testing: 2026-02-04 16:51:26  
Date of Next Testing: 2030-02-03 16:51:26  
Product Name: **titrant-component for volumetric Karl Fischer titration**  
CAS Number:  
Specifications & Purity: Water equivalent 3 mg H<sub>2</sub>O/mL,B  
Lot #: B2603611  
SKU #: **K301080**  
Version: **2**  
Analysis #: **640787**

Parameter	Limit Values	Result
pH	4-6	5.30
Titration strength(H <sub>2</sub> O)	>=3(mg/ml)	3.85750 mg/ml
Sensitivity	Pass	Pass



John Su  
QA & QC Manager