

Certificate of Analysis

Date of Testing: 2024-09-18 16:23:42
Date of Next Testing: 2026-09-08 16:23:42
Product Name: **one-component reagent for volumetric Karl Fischer titration (pyridine-free titrating agent)**
CAS Number:
Specifications & Purity: 3-5mg H₂O/ml, Sample for routine water content measurement
Lot #: I2412397
SKU #: **K492606**
Version: **1**
Analysis #: **407080**

| Parameter | Limit Values | Result |
|-----------|--------------|------------|
| Titration | 3-5 (mg/ml) | 3.72 mg/ml |



John Su
QA & QC Manager