

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2023-10-08 16:08:53
Date of Next Testing: 2025-10-07 16:08:53
Product Name: **Lauric acid**
CAS Number: 143-07-7
Specifications & Purity: Melting point standard
Lot #: A2206414
SKU #: **L119549**
Analysis: **282485**

Parameter	Limit Values	Result
Melting point(0.50°C/min)	44.19-44.51 (°C)	44.22 °C
Melting point(0.20°C/min)	44-44.38 (°C)	44.21 °C
Melting point(1.0°C/min)	44.53-44.95 (°C)	44.57 °C



John Su
QA & QC Manager