

## Certificate of Analysis

Date of Testing: 2024-11-15 13:30:25  
Date of Next Testing: 2026-11-15 13:30:25  
Product Name: **Lauric acid**  
CAS Number: 143-07-7  
Specifications & Purity: Moligand™, Melting point standard  
Lot #: D2117231  
SKU #: **L119549**  
Version: **0**  
Analysis #: **437222**

Parameter	Limit Values	Result
Melting point(0.50°C/min)	44.19-44.51 (°C)	44.20 °C
Melting point(0.20°C/min)	44-44.38 (°C)	44.10 °C
Melting point(1.0°C/min)	44.53-44.95 (°C)	44.60 °C



John Su  
QA & QC Manager