

Certificate of Analysis

Date of Testing: 2024-07-19 11:05:03
Date of Next Testing: 2026-07-19 11:05:03
Product Name: **Nickel Standard**
CAS Number: 7440-02-0(solution)
Specifications & Purity: 1000µg/ml in 1mol/L HNO₃
Lot #: G2410425
SKU #: **N115430**
Version: **0**
Analysis #: **391006**

Parameter	Limit Values	Result
Concentration	990-1010 (µg/ml)	1000.00 µg/ml
Appearance (N115430Form)	Liquid	Liquid
Appearance (N115430Color)	Green transparent	Green transparent



John Su
QA & QC Manager