

Certificate of Analysis

Date of Testing: 2025-04-21 17:33:58

Date of Next Testing: 2027-04-21 17:33:58

Product Name: **Nitrate Standard**

CAS Number:

Specifications & Purity: 1000µg/ml in Water (20°C)

Lot #: D2516524

SKU #: **N117308**

Version: **1**

Analysis #: **501314**

Storage Temperature: Protected from light, Room temperature

Parameter	Limit Values	Result
Concentration	990-1010 (mg/L)	1000.00 mg/L
Related expanded uncertainty(k=2)	0-4 (%)	0.30000 %
Appearance (N117308)	Colorless, transparent liquid	Colorless, transparent liquid

A handwritten signature in black ink that reads "John Su".

John Su
QA & QC Manager