

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

## Certificate of Analysis

Date of Testing: 2026-03-25 16:45:54

Date of Next Testing: 2027-03-25 16:45:54

Product Name: **Nitrate Reduction Test Reagent (Griess Reagent, without Zinc Reducer)**

CAS Number:

Specifications &amp; Purity: for microscopy, Suitable for Microbiology, Biological Stain, BioReagent

Lot #: ZJ26F0333465

SKU #: **N1510381-CG2×50ml**

Version: **1**

Analysis #: 029710

| Parameter                                                  | Limit Value                              | Result                                   |
|------------------------------------------------------------|------------------------------------------|------------------------------------------|
| Appearance(N1510381A)                                      | Colorless and transparent liquid         | Colorless and transparent liquid         |
| Appearance(N1510381B)                                      | Brown liquid                             | Brown liquid                             |
| Seal Integrity Testing                                     | Conform                                  | Conform                                  |
| Performance (Positive (Enterobacteriaceae, Veillonella))   | Red without zinc reducing agent          | Red without zinc reducing agent          |
| Performance(Negative)                                      | Red with zinc reducing agent             | Red with zinc reducing agent             |
| Performance (False negative (Pseudomonas, Serratia, etc.)) | No color change with zinc reducing agent | No color change with zinc reducing agent |



Sophia Cheng  
Quality Assurance