

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-07-03 18:25:48
 Date of Next Testing: 2027-06-23 18:25:48
 Product Name: **Niobium standard**
 CAS Number: 7440-03-1
 Specifications & Purity: 1000µg/ml, in 1.0 mol/L hydrofluoric acid and trace HNO₃
 Lot #: F2523597
 SKU #: **N489893**
 Version: **2**
 Analysis #: **500697**
 Storage Temperature: Room temperature

Parameter	Limit Values	Result
Concentration	990-1010 (mg/L)	1000.00000 mg/L
Related expanded uncertainty(k=2)	0-4 (%)	0.70000 %
Solvent(hydrofluoric acid and HNO ₃)	Conforms	Conforms
Appearance(N489893)	Colorless to Faint Yellow Transparent Liquid	Colorless to Faint Yellow Transparent Liquid



John Su
QA & QC Manager