

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-03-07 10:09:07

Date of Next Testing: 2027-03-07 10:09:07

Product Name: **pH Buffer Standard Solution (pH=4.01)**

CAS Number:

Specifications & Purity: pH=4.012 (25°C)

Lot #: C2507004

SKU # **P118675**

Version **1**

Analysis #: **482804**

Storage Temperature: Room temperature

Parameter	Limit Values	Result
pH	3.99-4.03	4.00400
Appearance (P118675)	Colorless clear liquid	Colorless clear liquid

A handwritten signature in black ink that reads "John Su".

John Su
QA & QC Manager