

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-05-23 10:08:41

Date of Next Testing: 2027-05-23 10:08:41

Product Name: **pH Buffer Standard Solution (pH=4.01)**

CAS Number:

Specifications & Purity: pH=4.012 (25°C)

Lot #: E2523007

SKU # **P118675**

Version **1**

Analysis #: **516662**

Storage Temperature: Room temperature

Parameter	Limit Values	Result
PH (25°C)	3.99-4.03	4.01000
Appearance (P118675)	Colorless clear liquid	Colorless clear liquid



John Su
QA & QC Manager