

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-12-15 18:58:36

Date of Next Testing: 2027-12-15 18:58:36

Product Name: **pH Buffer Standard Solution (pH=4.01)**

CAS Number:

Specifications & Purity: pH=4.012 (25°C)

Lot #: L2508650

SKU # **P118675**

Version **1**

Analysis #: **613987**

Storage Temperature: Room temperature

Parameter	Limit Values	Result
PH (25°C)	3.99-4.03	3.99700
Appearance (P118675)	Colorless clear liquid	Colorless clear liquid



John Su
QA & QC Manager