

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-12-17 12:23:02

Date of Next Testing: 2027-12-17 12:23:02

Product Name: **pH Buffer Standard Solution (pH=4.01)**

CAS Number:

Specifications & Purity: pH=4.012 (25°C)

Lot #: L2516634

SKU # **P118675**

Version **1**

Analysis #: **627539**

Storage Temperature: Room temperature

Parameter	Limit Values	Result
PH (25°C)	3.99-4.03	4.00100
Appearance (P118675)	Colorless clear liquid	Colorless clear liquid

A handwritten signature in black ink that reads "John Su".

John Su
QA & QC Manager