

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-01-16 08:47:38

Date of Next Testing: 2027-01-16 08:47:38

Product Name: **pH Buffer Standard Solution (pH=6.86)**

CAS Number:

Specifications & Purity: pH=6.863(25°C)

Lot #: A2513229

SKU # **P118676**

Version **1**

Analysis #: **463687**

Storage Temperature: Room temperature

Parameter	Limit Values	Result
pH	6.84-6.88	6.87000
Appearance (P118676)	Colorless clear liquid	Colorless clear liquid

A handwritten signature in black ink that reads "John Su".

John Su
QA & QC Manager