

Certificate of Analysis

Date of Testing: 2023-03-15 09:57:13
Date of Next Testing: 2025-03-14 09:57:13
Product Name: **pH Buffer Standard Solution (pH=6.86)**
CAS Number:
Specifications & Purity: pH=6.863(25°C)
Lot #: C2314001
SKU #: **P118676**
Analysis: **199305**

Parameter	Limit Values	Result
PH	6.84-6.88	6.87
Appearance (P118676)	Colorless clear liquid	Colorless clear liquid



John Su
QA & QC Manager