

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2023-01-31 13:48:09

Date of Next Testing: 2025-01-30 13:48:09

Product Name: **pH Buffer Standard Solution (pH=9.18)**

CAS Number:

Specifications & Purity: pH=9.184 (25°C)

Lot #: A2330008

SKU # **P118677**

Version **1**

Analysis #: **180960**

Storage Temperature: Room temperature

Parameter	Limit Values	Result
pH	9.164-9.204	9.17300
Appearance (P118677)	Colorless clear liquid	Colorless clear liquid



John Su
QA & QC Manager