

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

## Certificate of Analysis

Date of Testing: 2025-12-15 19:10:45

Date of Next Testing: 2027-12-15 19:10:45

Product Name: **pH Buffer Standard Solution (pH=9.18)**

CAS Number:

Specifications & Purity: pH=9.184 (25°C)

Lot #: L2510602

SKU # **P118677**

Version **1**

**Analysis #:** **615211**

Storage Temperature: Room temperature

| Parameter            | Limit Values           | Result                 |
|----------------------|------------------------|------------------------|
| PH (25°C)            | 9.164-9.204            | 9.17600                |
| Appearance (P118677) | Colorless clear liquid | Colorless clear liquid |



John Su  
QA & QC Manager