

Certificate of Analysis

Date of Testing: 2024-11-23 16:06:52

Date of Next Testing: 2026-11-23 16:06:52

Product Name: **pH Buffer Standard Solution (pH=6.86±0.02)**

CAS Number:

Specifications & Purity: pH=6.86±0.02 (25°C)

Lot #: B2307002

SKU # **P118679**

Version **0**

Analysis #: **440006**

Parameter	Limit Values	Result
PH	6.84-6.88	6.86
Appearance (P118679)	Colorless clear liquid	Colorless clear liquid



John Su
QA & QC Manager