

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-04-27 09:58:38

Date of Next Testing: 2027-04-27 09:58:38

Product Name: **pH Buffer Standard Solution (pH=6.86±0.02)**

CAS Number:

Specifications & Purity: pH=6.86±0.02 (25°C)

Lot #: B2309001

SKU # **P118679**

Version **1**

Analysis #: **503790**

Parameter	Limit Values	Result
PH (25°C)	6.84-6.88	6.86
Appearance (P118679)	Colorless clear liquid	Colorless clear liquid

A handwritten signature in black ink that reads "John Su".

John Su
QA & QC Manager