

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-04-16 13:41:14

Date of Next Testing: 2027-04-16 13:41:14

Product Name: **pH Buffer Standard Solution (pH=4.00±0.02)**

CAS Number:

Specifications & Purity: pH=4.00±0.02 (25°C)

Lot #: D2511401

SKU # **P118681**

Version **1**

Analysis #: **498453**

Parameter	Limit Values	Result
PH (25°C)	3.98-4.02	3.99
Appearance (P118681)	Colorless clear liquid	Colorless clear liquid

A handwritten signature in black ink that reads "John Su".

John Su
QA & QC Manager