

Certificate of Analysis

Date of Testing: 2024-07-01 11:57:19

Date of Next Testing: 2026-07-01 11:57:19

Product Name: **pH Buffer Standard Solution (pH=4.00±0.02)**

CAS Number:

Specifications & Purity: pH=4.00±0.02 (25°C)

Lot #: I2222431

SKU # **P118681**

Version **0**

Analysis #: **378352**

Parameter	Limit Values	Result
PH	3.98-4.02	4.00
Appearance (P118681)	Colorless clear liquid	Colorless clear liquid



John Su
QA & QC Manager