

Certificate of Analysis

Date of Testing: 2025-06-24 09:33:21

Date of Next Testing: 2027-06-24 09:33:21

Product Name: **pH Buffer Standard Solution (pH=10.00±0.02)**

CAS Number:

Specifications & Purity: pH=10.00±0.02 (25°C)

Lot #: I2312526

SKU # **P118683**

Version **0**

Analysis #: **533444**

Parameter	Limit Values	Result
PH	9.98-10.02	9.99
Appearance (P118683)	Colorless clear liquid	Colorless clear liquid



John Su
QA & QC Manager