

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2024-03-20 11:29:10
Date of Next Testing: 2026-03-20 11:29:10
Product Name: **Phenacetin**
CAS Number: 62-44-2
Specifications & Purity: Melting point standard
Lot #: F2214602
SKU #: **P119536**
Analysis: **333342**

Parameter	Limit Values	Result
Capillary melting point(0.2 Celsius/min)	134.67-134.95 (°C)	134.80 °C
Capillary melting point(0.5 Celsius/min)	134.93-135.25 (°C)	135.10 °C
Capillary melting point(1.0 Celsius/min)	135.24-135.6 (°C)	135.30 °C



John Su
QA & QC Manager