



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2026-04-22 08:45:53
Date of Next Testing: 2028-04-21 08:45:53
Product Name: **PCB No 52**
CAS Number: 35693-99-3
Specifications & Purity: 100 ug/mL in Isooctane
Lot #: D2615679
SKU #: **P128465**
Version: **1**
Analysis #: **696354**
Storage Temperature: Store at 2-8°C

Parameter	Limit Values	Result
Concentration	95-105 (mg/L)	101.00000 mg/L
Appearance(Colorless Transparent Liquid)	Pass	Pass
Record the entire process by video	Conform	Conforms

John Su
QA & QC Manager