



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-01-08 11:38:37
Date of Next Testing: 2029-01-07 11:38:37
Product Name: **pH Buffers Solutions(pH 1.00±0.02)**
CAS Number:
Specifications & Purity: Buffer Solution pH 1.00±0.02
Lot #: A2507015
SKU # **P196900**
Version **0**
Analysis #: **459011**

Parameter	Limit Values	Result
PH	0.98-1.02	0.99
Appearance (P196900)	Colorless clear liquid	Colorless clear liquid

John Su
QA & QC Manager