



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2022-12-27 16:29:31
Date of Next Testing: 2026-12-26 16:29:31
Product Name: pH Buffers Solutions(pH 1.00±0.02)
CAS Number:
Specifications & Purity: Buffer Solution pH 1.00±0.02
Lot #: L2227001
SKU # P196900
Analysis: 171712

Parameter	Limit Values	Result
PH	0.98-1.02	1.02
Appearance (P196900)	Colorless clear liquid	Colorless clear liquid

John Su
QA & QC Manager