



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2023-09-18 13:36:40
Date of Next Testing: 2027-09-17 13:36:40
Product Name: pH Buffers Solutions(pH 6.00±0.01)
CAS Number:
Specifications & Purity: Buffer Solution pH 6.00±0.01
Lot #: I2312527
SKU # P196905
Analysis: 276720

Parameter	Limit Values	Result
PH	5.99-6.01	6.00
Appearance (P196905)	Colorless clear liquid	Colorless clear liquid

John Su
QA & QC Manager