



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-06-12 16:15:45
Date of Next Testing: 2029-06-11 16:15:45
Product Name: **pH Buffers Solutions(pH 6.80±0.01)**
CAS Number:
Specifications & Purity: Buffer Solution pH 6.8±0.01
Lot #: G2113008
SKU # **P196906**
Version **0**
Analysis #: **528368**

Parameter	Limit Values	Result
PH	6.79-6.81	6.80
Appearance (P196906)	Colorless clear liquid	Colorless clear liquid

John Su
QA & QC Manager